



Certificate of Calibration

Reference: 10022-0700 Model: SDS 97D HF
S/N: 2005

To Whom It May Concern:

This is to certify that the above-referenced instrument was calibrated at **ENMET** Corporation

Date: 5-27-2008

As part of your Purchase Order:

The calibration was performed using **ENMET** procedures and gas generated by a Permeation Tube

Gas: Concentration: 5 ppm HF ALARM POINT 3.6, 9

Perm Tube #: 110-094-6040

Diffusion Rate: 3.948 ng/min. at 30 °C

We trust the above is sufficient for your purposes.

Sincerely
ENMET Corporation

A handwritten signature in red ink, appearing to read "O. Phuley", is written over the company name.

Quality Control Department

Gas and Vapor Detection Instruments for Industrial Health and Safety Applications
680 Fairfield Court, P.O. Box 979, Ann Arbor, MI 48106-0979, Phone: 734-761-1270, Fax: 734-761-3220
web: www.enmet.com e-mail: info@enmet.com



Certification of Conformance

Description: GS-24-DF-97D

Date: 5-27-08

Part Number: 04565-310

Serial Number: 461

ENMET certifies that the above part conforms to the specifications and requirements stated on the purchase order and drawings at time of shipment.

Sincerely,

ENMET Corporation

A handwritten signature in red ink, appearing to read "O. Kelly", is written over the printed name of the Quality Control coordinator.

Quality Control coordinator

